

CT Lower Extremity Guideline

Indications: Evaluate lesion, Fracture

Oral Contrast: None

IV Contrast: If indicated; Yes (See [MRG Contrast Media Administration Policy](#))

Positioning	Supine or Prone-feet or head first
FOV	Patient size
Axial Reconstruction: Thickness/ Interval/ Algorithm	(1) 1mm/1mm/Sharp (2) 1mm/1mm/Soft
Start Location	Above area of interest or fracture
End Location	Through area of interest or fracture
Gantry Angle	None
Reformations: Thickness/ Interval/ Algorithm	(1) Coronal 2mm/2mm/Sharp (2) Sagittal 2mm/2mm/Sharp
Phase of Respiration	N/A