

MRI Abdomen Appendix Guideline (62A)
CPT Code: 74181

Indications: RLQ pain during Pregnancy, Appendicitis

Contrast: No

Sequence	Plane	Slice Thickness	Gap	FOV	Notes
T2	Coronal	5mm	.5mm	350-450	
T2	Axial Upper	5mm	.5mm	350-450	
T2	Axial Lower	5mm	.5mm	350-450	
T1 Flash	Axial Upper	5mm	.5mm	350-450	
T1 Flash	Axial Lower	5mm	.5mm	350-450	
STIR	Axial Upper	5mm	.5mm	350-450	
STIR	Axial Lower	5mm	.5mm	350-450	
T1 Flash	Coronal	5mm	.5mm	350-450	
STIR	Coronal	5mm	.5mm	350-450	