

**MRI Prostate Pelvis With and Without Guideline (65)**
**CPT Code:** 72197

**Indications:** Staging of prostate CA, Elevated PSA, Prostate CA Screening

**Contrast:** Yes (see MRG MRI Contrast Administration Policy)

**Non Contrast:**

Sequence	Plane	Slice Thickness	Gap	FOV	Notes
T1	Axial	4mm	1mm	350-450	Entire Pelvis
T2 STIR	Axial	4mm	1mm	350-450	Entire Pelvis
T2	Sagittal	3mm	0	140	
T2	Axial Oblique	3mm	0	140	
T2	Coronal Oblique	3mm	0	140	
T1	Axial Oblique	3mm	0	140	FatSat
DWI	Axial	3mm	0	220	B value 50, 800
DWI	Axial	3mm	0	220	B Value 1400

**\*INJECT CONTRAST AFTER 5th MEASURE**
**Contrast:**

Sequence	Plane	Slice Thickness	Gap	FOV	Notes
T1 TWIST	Axial Obl Dyn	3mm	0	200	
T1	Axial Oblique	3mm	0	140	FatSat
T1	Axial	3mm	0	350-450	FatSat, Entire Pelvis